

## Addendum to MiniScan EZ User's Manual

---

### Description



This CMR modifies the standard MiniScan EZ calibration tiles to a raised position slightly above flush.

### Installation

No special installation required.

### Operation

Place the calibration tiles at the port of the MiniScan EZ instrument when prompted to do so during standardization and when performing diagnostics.